



Issued to: **Cordes Courant**
36, boulevard de l'industrie,
ZI Ecoouflant 49000 Angers
France

Notified Body 0321

SATRA Client: P0405

EC Type-Examination Certificate

Number 5231 Issue 1 Extension 1

This is to certify that the product group reference RMQCMOKA comprising the following products

Product Reference	Description
RMQCMOKA	Aluminium karabiner with self-locking gate triple action gate Static strength of 20kN Gate opening of 19mm

Classification:

EN 362:2004 Class B

EN 12275:1998 Type B

Technical reports

SATRA: SPC0190735/1101/1 Issue 2

has been subject to an EC Type-examination in accordance with Article 10 of the PPE Directive (89/686/EEC) and has been shown to satisfy the relevant provisions of this Directive for the complex category through:

- i Testing to the following standard: EN 362:2004 and EN 12275:1998
- ii Examination of the relevant technical documentation.

You are therefore licensed to mark the product(s) listed above in accordance with Article 13 of Directive (89/686/EEC) and any relevant amending Directives once you have drawn up an EC declaration of product conformity. Please note that:

1. Full details of the certification and product are contained in the manufacturer's technical file
2. This certificate is only valid if embossed with the text – "SATRA – European Notified Body 0321"
3. This certificate is issued subject to the conditions on the reverse side of this certificate
4. CE marking of production is also reliant on current compliance with Directive 89/686/EEC Article 11

Signed:  (G Graham)

Signed: 

Date: 11th December 2014

On behalf of SATRA